

College of American Pathologists (CAP) GH5 Survey Data:

(updated 12/17)

The American Diabetes Association (ADA) recommends that “The A1C test should be performed using a method that is certified by the NGSP”. All laboratories performing HbA1c testing should participate in a fresh sample proficiency testing survey such as the College of American Pathologists (CAP). CAP GH5 data for the **third** survey of 2017 are summarized below. The NGSP target or reference values are based on replicate analyses using eight NGSP certified secondary reference methods.

Commentary by R. Little, Ph.D., NGSP Network Coordinator for the NGSP Steering Committee

Beginning in 2015 there are two CAP programs for HbA1c proficiency testing using fresh whole blood samples - GH2 and GH5. GH2 samples are shipped twice a year with three samples in each mailing as before. GH5 are shipped three times a year with five samples in each mailing. The three samples in each of the two GH2 mailings are also included in two of the GH5 mailings. Therefore the NGSP follows the three GH5 surveys which include all the samples used for both surveys.

In 2017, based on data from the GH5-C survey:

- Bias from the NGSP target and variability ($\pm 2SD$) are shown in Table 1 and in figure 1 (ordered by HbA1c level in figure) for each method. The shaded rectangle (fig 1) reflects the current CAP acceptance limit of ± 6 .
- Method-specific biases > 0.30 (shaded cells, table) were seen with two methods (Roche Cobas Integra 400 and Siemens DCA 2000/2000+, each for 1 sample).
- Method-specific, between-laboratory CV's ranged from 1.1% to 4.3%. The Siemens Dimension RxL had CVs over 3.5% for two samples and the Beckman AU had a CV over 3.5% for one sample. There were five methods with CVs $\leq 2\%$ for 5/5 samples– the Abbott Architect c enzymatic, Bio-Rad D-100, Roche cobas c513, Sebia Capillarys 2 and Tosoh G8. Arkray Adams HA-8180 had CVs $\leq 2\%$ for 3/3 samples. Approximately 72% of laboratories are using methods with between-lab CVs $\leq 3\%$ at all five HbA1c levels; approximately 97% of laboratories are using methods with CVs $\leq 3.5\%$ at all five HbA1c levels.
- The current pass limit for the GH5 survey is $\pm 6\%$. The overall pass rates for this survey were 96.8%, 96.9%, 97.2%, 96.4% and 97.5% for GH5-11 through 15, respectively. For individual methods, the lowest pass rate was 72.7% and the highest was 100%. As expected, methods with small bias and low CVs will have the highest pass rates and, conversely, methods with large bias and/or high CVs will have the lowest pass rates.
- The overall CVs for the last 19 surveys are shown in Table 2. CVs were $\leq 3.0\%$ for all 5 samples in the current survey.


NOTE: The NGSP certification evaluates agreement of each method at the manufacturing site using one lot of reagents and calibrators, one instrument, and one application under optimal conditions. CAP precision reflects between-laboratory reproducibility, often with more than one lot of reagents and calibrators, and sometimes with different instruments (e.g. Cobas Integra 400 & Cobas Integra 800) and/or different applications (e.g. Cobas Integra hemolysate or whole blood application). In addition, if changes were made in the method just prior to NGSP certification, it is possible that not all participating laboratories in the field would have made the change at the time of the CAP survey. For these reasons, it is important that laboratories review not only the certification status of HbA1c methods but also their performance in the CAP survey over time (a good indication of field performance) when selecting or evaluating HbA1c assay methods.

TABLE 1: 2017 GH5-C (fresh pooled samples)

		GH5-11			GH5-12			GH5-13			GH5-14			GH5-15		
NGSP %HbA1c Reference Value (95% CI)		7.25 (7.18-7.31)			6.81 (6.74-6.88)			8.65 (8.58-8.73)			9.50 (9.43-9.57)			5.45 (5.38-5.52)		
	no. labs	Mean %HbA1c	Mean bias	% CV	Mean %HbA1c	Mean bias	% CV	Mean %HbA1c	Mean bias	% CV	Mean %HbA1c	Mean bias	% CV	Mean %HbA1c	Mean bias	% CV
Abbott Architect c (enzymatic)	189	7.14	-0.11	1.3	6.67	-0.14	1.4	8.60	-0.05	1.2	9.47	-0.03	1.2	5.29	-0.16	1.3
Alere Afinion	113	7.15	-0.10	2	6.74	-0.07	2.0	8.78	0.13	2.1	9.23	-0.27	2.2	5.34	-0.11	1.5
Arkray Adams HA-8180 series	19	7.35	0.10	1.1	6.82	0.01	1.6	8.70	0.05	1.4						
Beckman AU	91	7.26	0.01	3	6.83	0.02	3.6	8.55	-0.10	2.3	9.43	-0.07	2.3	5.43	-0.02	2.9
Beckman UniCel DxC	118	7.33	0.08	3.1	6.84	0.03	3.0	8.77	0.12	2.6	9.78	0.28	2.7	5.45	0.00	2.3
Bio-Rad D-10	156	7.36	0.11	2.3	6.90	0.09	2.4	8.71	0.06	2.1	9.65	0.15	2.4	5.48	0.03	2.7
Bio-Rad D-100	41	7.19	-0.06	1.7	6.76	-0.05	1.7	8.49	-0.16	1.3	9.32	-0.18	2.0	5.42	-0.03	1.7
Bio-Rad VII	39	7.32	0.07	2.4	6.92	0.11	2.5	8.66	0.01	2.4	9.56	0.06	2.6	5.47	0.02	2.7
Bio-Rad VII Turbo	67	7.31	0.06	2.5	6.85	0.04	2.3	8.72	0.07	2.8	9.66	0.16	2.9	5.41	-0.04	2.5
Bio-Rad VII Turbo 2.0	152	7.29	0.04	2	6.86	0.05	2.0	8.58	-0.07	1.8	9.54	0.04	1.9	5.49	0.04	2.2
Roche Cobas c311	26	7.10	-0.15	1.8	6.56	-0.25	2.2	8.64	-0.01	1.6	9.55	0.05	1.4	5.16	-0.29	2.6
Roche Cobas c500 series	406	7.16	-0.09	2.1	6.67	-0.14	2.2	8.58	-0.07	2.2	9.48	-0.02	2.3	5.31	-0.14	2.2
Roche Cobas c513	34	7.24	-0.01	1.5	6.78	-0.03	1.4	8.69	0.04	1.6	9.67	0.17	1.5	5.42	-0.03	1.3
Roche Cobas Integra 400	50	7.22	-0.03	2.8	6.73	-0.08	2.7	8.77	0.12	2.8	9.94	0.44	3.5	5.51	0.06	3.4
Roche Cobas Integra 800	38	7.22	-0.03	1.7	6.79	-0.02	2.6	8.69	0.04	2.4	9.60	0.10	2.8	5.55	0.10	2.7
Sebia Capillarys 2	53	7.10	-0.15	1.5	6.70	-0.11	1.9	8.50	-0.15	1.3	9.33	-0.17	1.6	5.34	-0.11	2.0
Siemens Advia	16	7.29	0.04	2.8	6.79	-0.02	2.8	8.59	-0.06	2.3	9.31	-0.19	2.3	5.45	0.00	2.8
Siemens DCA 2000/2000+	10	7.43	0.18	3	7.08	0.27	3.2	8.96	0.31	3.1						
Siemens DCA Vantage	435	7.31	0.06	2.8	6.93	0.12	2.8	8.73	0.08	2.9	9.74	0.24	3.5	5.48	0.03	2.3
Siemens Dimension ExL	206	7.29	0.04	2.5	6.96	0.15	3.1	8.53	-0.12	2.9	9.49	-0.01	2.5	5.51	0.06	3.1
Siemens Dimension RxL	15	7.29	0.04	4.3	7.05	0.24	3.0	8.59	-0.06	1.7	9.44	-0.06	2.5	5.54	0.09	4.2
Siemens Dimension Vista	300	7.45	0.20	2.3	7.01	0.20	2.5	8.46	-0.19	2.1	9.28	-0.22	2.1	5.54	0.09	2.8
Siemens Dimension Xpand	26	7.27	0.02	2.8	6.90	0.09	3.1	8.49	-0.16	3.1	9.43	-0.07	3.4	5.56	0.11	3.2
Tosoh G8 Auto HPLC	378	7.35	0.10	1.4	6.99	0.18	1.4	8.72	0.07	1.3	9.68	0.18	1.2	5.54	0.09	1.4
Trinity Biotech Premier	79	7.28	0.03	2.3	6.79	-0.02	2.3	8.92	0.27	2.3	9.58	0.08	2.1	5.55	0.10	2.4
(Ortho Clin Diag) Vitros 5,1 FS, 4600, 5600	194	7.28	0.03	2.6	6.83	0.02	2.4	8.66	0.01	2.7	9.60	0.10	3.0	5.42	-0.03	2.1

Gray shading indicates bias > 0.3% HbA1c or CV > 3.5% Note: these are arbitrary limits chosen to highlight methods with the highest bias and CV.

Figure 1: Bias and Variability from the NGSP Target







Table 2: Overall Variability for 2010-2017 for all GH participants

Mailing	Sample#	# of labs	Target	All method mean	S.D.	C.V.
A-2010	01	2573	5.9	6.03	0.23	3.9
	02	2566	9.8	9.73	0.39	4.0
	03	2581	7.4	7.43	0.31	4.2
B-2010	04	2693	5.2	5.34	0.21	4.0
	05	2691	8.7	8.67	0.33	3.8
	06	2685	6.3	6.37	0.23	3.5
A-2011	01	2652	8.5	8.58	0.28	3.2
	02	2645	5.4	5.52	0.20	3.5
	03	2649	6.4	6.51	0.21	3.2
B-2011	04	2877	6.3	6.36	0.24	3.8
	05	2872	7.6	7.69	0.29	3.8
	06	2871	9.2	9.28	0.34	3.7
A 2012	01	3298	5.6	5.62	0.20	3.5
	02	3316	9.4	9.44	0.37	3.9
	03	3301	7.2	7.28	0.29	3.9
B2012 (HbAS)	04	3222	5.4	5.51	0.21	3.9
	05	3208	8.3	8.31	0.31	3.7
	06	3172	5.65	5.75	0.32	5.6
A 2013	01	2816	7.1	7.12	0.25	3.5
	02	2829	9.3	9.39	0.31	3.3
	03	2840	6.1	6.13	0.24	3.9
B2013	04	2912	8.1	8.04	0.31	3.8
	05	2907	5.3	5.33	0.20	3.8
	06	2908	6.4	6.17	0.24	3.9
A2014	01	3277	6.5	6.60	0.25	3.8
	02	3267	7.0	7.09	0.27	3.8
	03	3253	9.7	9.72	0.33	3.4
B2014	04	3278	6.58	6.64	0.23	3.5
	05	3273	8.39	8.45	0.30	3.6
	06	3266	5.65	5.67	0.21	3.6
A2015	01	3237	6.79	6.82	0.25	3.6
	02	3246	10.28	10.19	0.36	3.5
	03	3252	6.82	6.82	0.25	3.6
	04	2365	8.63	8.63	0.30	3.4
	05	2362	5.32	5.36	0.18	3.4
B2015	06	2379	5.84	5.87	0.2	3.5
	07	2392	11.71	11.68	0.44	3.8
	08	2402	9.53	9.5	0.33	3.5
	09	2386	5.04	5.08	0.17	3.4
	10	2403	7.38	7.35	0.26	3.5
C2015	11	3284	11.69	11.68	0.47	4.1
	12	3285	5.93	5.95	0.19	3.3
	13	3286	5.17	5.2	0.17	3.3
	14	2410	8.14	8.12	0.24	2.9
	15	2408	9.3	9.25	0.29	3.2
A2016	01	3358	5.32	5.33	0.16	3.1

	02	3365	9.17	9.21	0.28	3.0
	03	3357	5.31	5.33	0.16	3.1
	04	2425	12.03	12.12	0.40	3.3
	05	2419	5.94	5.96	0.16	2.8
B2016	06	2433	5.27	5.27	0.15	2.8
	07	2427	10.59	10.55	0.33	3.1
	08	2440	6.20	6.17	0.18	2.9
	09	2428	12.23	12.21	0.44	3.6
	10	2443	7.51	7.52	0.20	2.7
C2016	11	3377	9.11	9.08	0.24	2.6
	12	3402	6.01	5.99	0.16	2.6
	13	3372	11.71	11.69	0.39	3.4
	14	2432	5.02	5.02	0.16	3.2
	15	2442	7.58	7.58	0.19	2.5
A2017	01	3418	6.41	6.45	0.2	3.1
	02	3393	9.53	9.56	0.28	2.9
	03	3409	5.34	5.35	0.17	3.1
	04	2461	8.51	8.58	0.23	2.7
	05	2460	7.25	7.33	0.19	2.6
B2017	06	2446	7.42	7.49	0.22	3.0
	07	2445	5.2	5.22	0.19	3.6
	08	2450	8.31	8.33	0.21	2.5
	09	2443	10.33	10.38	0.31	3.0
	10	2457	5.87	5.91	0.19	3.2
C2017	11	3331	7.25	7.28	0.19	2.6
	12	3339	6.81	6.85	0.2	3.0
	13	3344	8.65	8.64	0.23	2.6
	14	2431	9.5	9.54	0.26	2.8
	15	2419	5.45	5.45	0.16	2.9

CVs **below** 3.5% are highlighted in pink

CVs **below** 3.0% are highlighted in blue